



KAPA SYBR® FAST qPCR Kit Master Mix (2X) Universal

KR0389 – v9.13

Product Description

KAPA SYBR® FAST qPCR Master Mix is designed for high performance real-time PCR. The kit contains a novel DNA polymerase engineered via a process of molecular evolution. The result is a unique enzyme, specifically designed for qPCR using SYBR® Green I dye chemistry.

KAPA SYBR® DNA Polymerase has been engineered to perform optimally in stringent real-time quantitative PCR (qPCR) reaction conditions, exhibiting dramatic improvements in signal-to-noise ratio (fluorescence), quantification cycle (C_q), linearity, and sensitivity. The KAPA SYBR® DNA Polymerase and proprietary buffer system improves the amplification efficiency of difficult targets, including both GC- and AT-rich templates.

KAPA SYBR® FAST qPCR Master Mix (2X) Universal is a ready-to-use cocktail containing all components except primers and template, for the amplification and detection of DNA in qPCR. The KAPA SYBR® FAST qPCR Kit is supplied as a 2X Master Mix with integrated antibody-mediated hot start, SYBR® Green I fluorescent dye, $MgCl_2$, dNTPs and stabilizers. ROX reference dye is not included in the 2X Master Mix but is supplied separately.

Product Applications

KAPA SYBR® FAST qPCR Kits are ideally suited for:

- Gene expression analysis
- Low copy gene detection
- Absolute quantification of NGS libraries.

Product Specifications

Shipping and Storage

Upon arrival, store kit components protected from light at $-20\text{ }^{\circ}\text{C}$ in a constant-temperature freezer. When stored under these conditions and handled correctly, the Master Mix (2X) will retain full activity until the expiry date indicated on the kit label.

Handling

Minimize exposure of the Master Mix (2X) to direct light. Exposure to direct light for an extended period of time may result in loss of fluorescent signal intensity. Always ensure that the product has been fully thawed and mixed before use.

Quality Control

KAPA SYBR® FAST qPCR Master Mix (2X) is free of contaminating DNase and RNase. It is functionally tested by generating a standard curve with human genomic DNA as template, with a dynamic range of 5 orders of magnitude, a reaction efficiency of 90 – 110% and a correlation coefficient >0.99 .

Effective date: October 2013

Kit Codes and Components

KK4600 100 x 20 μl reactions	KAPA SYBR® FAST Master Mix (2X) Universal qPCR Master Mix (2X) – 1 x 1 ml ROX Reference Dye High (50X) – 1 x 200 μl ROX Reference Dye Low (50X) – 1 x 200 μl
KK4601 500 x 20 μl reactions	KAPA SYBR® FAST Master Mix (2X) Universal qPCR Master Mix (2X) – 1 x 5 ml ROX Reference Dye High (50X) – 1 x 200 μl ROX Reference Dye Low (50X) – 1 x 200 μl
KK4602 1000 x 20 μl reactions	KAPA SYBR® FAST Master Mix (2X) Universal qPCR Master Mix (2X) – 2 x 5 ml ROX Reference Dye High (50X) – 2 x 200 μl ROX Reference Dye Low (50X) – 2 x 200 μl

Quick Notes

- This kit contains an engineered enzyme optimized for qPCR using SYBR® Green I dye chemistry.
- The 2X Master Mix contains a proprietary buffer. Together with the novel enzyme, this improves amplification efficiency of both GC- and AT-rich targets.
- 20 sec initial denaturation at $95\text{ }^{\circ}\text{C}$ is sufficient for enzyme activation. When working with complex templates, an initial denaturation of 3 min is recommended.
- For two-step cycling, use 20 sec combined annealing/extension/data acquisition.
- Do not exceed 25 μl reaction volumes.

Instrument	ROX Reference Dye
Applied Biosystems® 5700, 7000, 7300, 7700, 7900HT, StepOne™, and StepOnePlus™	ROX High
Applied Biosystems® 7500, ViiA™7, QuantStudio™ 12K Flex, Agilent Mx3000P™, Mx3005P™, and Mx4000™	ROX Low
Rotor-Gene™, DNA Engine Opticon™, Opticon™ 2, Chromo 4™ Real-Time Detector, Mastercycler® ep realplex, Smart Cycler®, Roche LightCycler® 480, Roche LightCycler® Nano, Bio-Rad CFX96, and Illumina Eco™	No ROX

KAPA SYBR® FAST qPCR Protocol

Any existing qPCR assay performed efficiently using standard cycling conditions may be converted to a fast qPCR assay with KAPA SYBR® FAST qPCR Kits. Typically, minimal re-optimization of reaction parameters is required.

Step 1: Prepare the qPCR master mix

- Ensure all reaction components are properly thawed and mixed.
- Prepare a PCR master mix containing the appropriate volume of all reaction components common to all or a subset of reactions to be performed.
- Calculate the required volume of each component based on the following table:

Component	20 µl rxn ¹	Final conc.
PCR-grade water	Up to 20 µl	N/A
2X KAPA SYBR® FAST qPCR Master Mix ² Universal	10 µl	1X
10 µM Forward Primer	0.4 µl	200 nM
10 µM Reverse Primer	0.4 µl	200 nM
Template DNA ³	As required	<20 ng
50X ROX High/Low (As required) ⁴	0.4 µl	1X

¹ Reaction volumes may be adjusted between 5 – 25 µl. Reaction volumes >25 µl are not recommended.

² KAPA SYBR® FAST qPCR Master Mix contains MgCl₂ at a final concentration of 2.5 mM.

³ Do not exceed 20 ng per 20 µl reaction. For more information, refer to Important Parameters: Template.

⁴ The use of ROX dye is necessary for all Applied Biosystems® instruments and is optional for the Agilent Mx3000P™, Mx3005P™, and Mx4000™ cyclers. Bio-Rad/MJ Research, Cepheid, Corbett/Qiagen, Eppendorf, Illumina and Roche instruments do not require ROX dye.

Step 2: Set up individual reactions

- Transfer the appropriate volumes of qPCR master mix, template and primers to each well of a PCR tube/plate.
- Cap or seal the reaction tube/plate and centrifuge briefly.

Step 3: Run the qPCR

- If applicable, select fast mode on the instrument.
- Confirm that qPCR protocol to be used conforms to the following parameters:

Step	Temperature	Duration	Cycles
Enzyme activation	95 °C	3 min ¹	Hold
Denaturation	95 °C	1 – 3 sec	40
Annealing/ extension/ data acquisition ³	60 °C	≥ 20 sec ²	
Dissociation	According to instrument guidelines		

¹ 20 sec at 95 °C is sufficient time for enzyme activation; however, optimal denaturation of complex targets may require up to 3 min denaturation.

² Select shortest time possible for instrument, but not less than 20 sec.

³ For 3-step cycling protocols, anneal at optimal annealing temperature for 20 sec followed by the minimum time required for data acquisition at 72 °C according to instrument guidelines.

Note: The above cycling parameters are not optimal for qPCR-based quantification of next-generation sequencing libraries. Please refer to the protocol in the KAPA Library Quantification Kit Technical Data Sheet.

Step 4: Analyze the results

- Melt curve analysis should be performed to identify the presence of primer-dimers and analyze the specificity of the reaction. Program your thermocycler according to the instrument guidelines.
- Data analysis is dependent on experimental design. Refer to your instrument guidelines for more information on how to perform the appropriate data analysis.

Important Parameters

Template

High concentrations of template may increase background fluorescence and reduce linearity of standard curves. For optimal quantitative results, use up to 20 ng of genomic DNA or plasmid DNA per 20 µl reaction (for smaller volumes, the amount of template should be decreased proportionally). For two-step RT-PCR, use either undiluted or diluted cDNA generated from up to 1 µg of total RNA. The volume of the cDNA (reverse transcription reaction product) should not exceed 10% of the final PCR volume (e.g. for a 20 µl qPCR reaction, use up to 2.0 µl of undiluted cDNA).

Important Parameters (continued)

Primers

Careful primer design and purification (HPLC-purified primers are recommended) will minimize loss in sensitivity due to nonspecific amplification. This effect becomes more prominent at low target concentrations. To maximize the sensitivity of the assay, use the lowest primer concentration determined not to compromise reaction efficiency (50 – 400 nM of each primer). For optimal results, design primers that amplify PCR products 60 – 400 bp in length. Use appropriate primer design software to design primers with a melting temperature (T_m) of approximately 60 °C, to take advantage of two-step cycling. If performing qRT-PCR, we recommend designing primers specifically for amplification of cDNA derived from mRNA. This prevents amplification of contaminating genomic DNA and inaccurate quantification of mRNA.

KAPA SYBR® DNA Polymerase

KAPA SYBR® DNA Polymerase is an engineered version of *Taq* DNA polymerase, designed specifically for real-time PCR using SYBR Green I chemistry. KAPA SYBR® DNA Polymerase displays no enzymatic activity at ambient temperature. This prevents the formation of misprimed products and primer-dimers during reaction setup, prior to the first denaturation step, and results in high PCR specificity and accurate quantification. The enzyme is activated during the initial denaturation step of the PCR. The activation of the enzyme is complete after 20 sec; however, complex targets may require up to 3 min for optimal denaturation. The hot start feature obviates the need to cool reactions during setup.

ROX Reference Dye

For certain real-time cyclers, the presence of ROX reference dye compensates for non-PCR-related variations in fluorescence detection. The fluorescence level of ROX reference dye does not change significantly during the course of real-time PCR, but provides a stable baseline against which PCR-related fluorescent signals are normalized. Thus, ROX dye compensates for differences in fluorescence detection between wells due to slight variations in reaction volume or differences in well position. The presence of ROX dye in the master mix does not interfere with real-time PCR on any instrument, since the dye is not involved in the reaction and has an emission spectrum different from that of SYBR® Green I.

SYBR® Green I

KAPA SYBR® FAST qPCR Master Mix (2X) contains an elevated, optimized concentration of the fluorescent dye, SYBR® Green I. High signal intensities are achieved as a result of increased tolerance to high concentrations of SYBR® Green I by the engineered KAPA SYBR® DNA Polymerase. SYBR® Green I binds all double-stranded DNA molecules, emitting a fluorescent signal on binding.

Magnesium chloride

The KAPA SYBR® FAST qPCR Master Mix contains an optimized $MgCl_2$ concentration. It is highly unlikely that additional $MgCl_2$ will improve reaction efficiency or specificity.

Note to Purchaser: Limited License

Certain applications of this product are covered by patents issued to parties other than Kapa Biosystems and applicable in certain countries. Purchase of this product does not include a license to perform any such applications. Users of this product may therefore be required to obtain a patent license depending upon the particular application and country in which the product is used.

Use of this product is covered by one or more of the following US patents and corresponding patent claims outside the US: 5,994,056, 6,171,785, and 5,928,907 (claim numbers 12-24, 27-28). The purchase of this product includes a limited, non-transferable immunity from suit under the foregoing patent claims for using only this amount of product for the purchaser's own internal research. No right under any other patent claim (such as apparatus or system claims in US Patent No. 6,814,934) and no right to perform commercial services of any kind, including without limitation reporting the results of purchaser's activities for a fee or other commercial consideration, is conveyed expressly, by implication, or by estoppel. This product is for research use only. Diagnostic uses under Roche patents require a separate license from Roche. Further information on purchasing licenses may be obtained by contacting the Director of Licensing, Applied Biosystems, 850 Lincoln Centre Drive, Foster City, California 94404, USA.

This product is provided under an agreement between Life Technologies Corporation, and Kapa Biosystems Inc., and the manufacture, use, sale or import of this product is subject to one or more of U.S. Patent Nos. 5,436,134; 5,658,751 and corresponding international equivalents, owned by Life Technologies Corporation. The purchase of this product conveys to the buyer the non-transferable right to use the purchased amount of the product and components of the product in research conducted by the buyer, where such research does not include testing, analysis or screening services for any third party in return for compensation on a per test basis. The buyer cannot sell or otherwise transfer (a) this product (b) its components or (c) materials made using this product or its components to a third party or otherwise use this product or its components or materials made using this product or its components for Commercial Purposes. Commercial Purposes means any activity by a party for consideration and may include, but is not limited to: (1) use of the product or its components in manufacturing; (2) use of the product or its components to provide a service, information, or data; (3) use of the product or its components for therapeutic, diagnostic or prophylactic purposes; or (4) resale of the product or its components, whether or not such product or its components are resold for use in research. For information on purchasing a license to this product for purposes other than research, contact Life Technologies Corporation, 5791 Van Allen Way, Carlsbad, CA 92008 USA or outlicensing@lifetech.com.

Licensed under U.S. Patent nos. 5,338,671 and 5,587,287 and corresponding patents in other countries.

The purchase of this product includes a limited, non-transferable license under specific claims of U.S. Patent Nos. 6,174,670, 6,569,627 and 5,871,908, owned by the University of Utah Research Foundation or Evotec Biosystems GmbH and licensed to Idaho Technology, Inc. and Roche Diagnostics GmbH, to use only the enclosed amount of product according to the specified protocols. No right is conveyed, expressly, by implication, or by estoppel, to use any instrument or system under any claim of U.S. Patent Nos. 6,174,670, 6,569,627 and 5,871,908, other than for the amount of product contained herein.

Applied Biosystems®, SYBR®, PRISM®, GeneAmp®, ViiA™, QuantStudio™, StepOne™, and StepOnePlus™ are trademarks or registered trademarks of Life Technologies Corporation. iCycler®, Mx3000P™, Mx3005™, Mx4000™, Rotor-Gene™, DNA Engine Opticon™, Chromo 4™, LightCycler®, Smart Cycler®, and Eco™ are trademarks or registered trademarks of their respective companies.

Troubleshooting

Symptoms	Possible Causes	Solutions
<p>Positive signal in no-template control (NTC)</p> <p>Presence of secondary, nonspecific peak in melt curve of sample</p>	<p>Reasons for a positive signal in a NTC and/or nonspecific amplification are often due to multiple factors that include:</p> <ol style="list-style-type: none"> 1. Contamination 2. Primer-dimer formation due to: <ul style="list-style-type: none"> • Incorrect cycling protocol • Sub-optimal primer annealing temperature (often due to differences between qPCR systems) • Primer and/or template degradation. Primers should be stored and diluted in 10 mM Tris-HCl (pH 8.0 – 8.5), not in PCR-grade water) • Sub-optimal primer synthesis • Sub-optimal primer design 	<p>Perform melt curve analysis (or run qPCR products on a gel) to determine if the product is specific or nonspecific (primer-dimer).</p> <p>If the NTC contains a specific product, the assay is contaminated:</p> <ul style="list-style-type: none"> • Discard all reagents, clean all pipettes and surfaces and prepare fresh stocks of primer, etc. <p>Note: The increased sensitivity of KAPA SYBR® FAST qPCR Kits may result in the detection of low levels of contamination in assays considered contaminant-free when using competitor kits containing wild-type <i>Taq</i> DNA polymerase.</p> <p>If the NTC and/or sample contains nonspecific product, assay optimization may be required:</p> <ul style="list-style-type: none"> • 30 sec combined annealing/extension time is recommended for most assays. Longer times may result in nonspecific amplification. • Increase the combined annealing/extension temperature in increments of 3 °C. • Decrease primer concentration. • Resynthesize or redesign primers. HPLC purified primers are recommended for low copy number detection, and often results in reduced primer-dimer formation.
<p>Low fluorescence intensity</p>	<p>Incorrect handling</p> <p>Incorrect concentration of ROX reference dye</p>	<p>SYBR® Green I dye is light sensitive; avoid exposure to light and repeated freeze-thaw cycles. Always thaw and mix solutions thoroughly before use.</p> <p>If the incorrect concentration of ROX reference dye is added to the master mix, the normalized signal may be lower than expected (if too much ROX has been added), or higher than expected (if too little ROX has been added). If using ABI instrumentation, analysis of the raw signal can always be performed with the ROX filter switched off.</p>
<p>Melting temperature of specific product is different from competitor kit</p>	<p>Differences in the buffer composition (e.g. salt concentration) of qPCR master mixes</p>	<p>Differences in master mix formulation may affect the melting temperature of the product slightly. A particular DNA fragment will melt at a higher temperature in a reaction buffer containing a higher salt concentration.</p>

For technical support please contact:
support@kapabiosystems.com

Boston, Massachusetts, United States
Int. Tel: +1 781 497 2933 Tel: +1 855 KAPA BIO
Fax: +1 781 497 2934
Email: sales@kapabiosystems.com